



精准检测  
PRECISE TESTING

## VERIFICATION OF CONFORMITY

**No.** : PTC23031305301S-LD01  
**Applicant** : Nanchang Weiwu Technology Co., Ltd.  
**Address** : Huaxing Cultural Plaza, Honggutan District, Nanchang City  
**Manufacturer** : Nanchang Weiwu Technology Co., Ltd.  
**Address** : Huaxing Cultural Plaza, Honggutan District, Nanchang City  
**Factory** : Nanchang Weiwu Technology Co., Ltd.  
**Address** : Huaxing Cultural Plaza, Honggutan District, Nanchang City  
**Product** : Portable oxygenator  
**Model** : L100  
**Ratings** : Charging: 12.6V 2A, Working: 12V , Rated Power: 15W (Adapter: input: 100-240V~ 50/60Hz 1A Max Output: 12.6V 2A)

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

### Low voltage (LVD) Directive - 2014/35/EU

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s)	Test Report(s) Number
EN 60335- 1:2012 + A11:2014 + A13:2017 + A1:2019 + A14:2019 + A2:2019 + A15:2021 EN 62233:2008	PTC23031305301S-LD01

This verification is part of the full test report(s) and should be read in conjunction with it. This verification is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Precise Testing & Certification (Guangdong) Co., Ltd., this verification is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable.



*Jacky Ou*

Jacky Ou  
Manager  
Date: 2023-11-21